



INFLATABLE PURGE RING:
 DEFLATED CONDITION: 48.75 INCH MAXIMUM OUTSIDE DIAMETER
 INFLATED CONDITION: 48.75 TO 49.125 INCH OUTSIDE DIAMETER
 MAXIMUM PRESSURE: 5 P.S.I.G.

| | |
|----------------------------|--|
| Date Received: 11/10/95 | LIGO-D950081-02-B Contractor # CBI ER-123, Rev. 2 |
| TIM #: | LIGO- |
| Date: / / | Approved. |
| TDM #: | LIGO- |
| Date: / / | Approved. |
| Gen'l Transmittal | LIGO- |
| Date: / / | Approved: |

APPROVED

M. Jellalain 11/10/95
CBI DATE

J. Jones 11/10/95
CALTECH DATE

SECTION A

| | |
|--|-------------------------------|
| Supplier's/Purchaser's No. | |
| CBI | |
| He MASS SPECTROMETER TEST ARRANGEMENT LIGO PROJECT | |
| Customer's No. | Contract No. |
| By <i>DR</i> Chkd <i>DR</i> Date 11-3-93 | 930212ER |
| J.L. RINGELSTETTER Engineering Supervisor | Dwg. ER-123 Rev. 2 Sheet - |
| LIGO-D950081-02-B | |

CAD FILE 302123.DGN

INDICATES CHANGE FROM PREVIOUS ISSUE