



AGM CONTAINER CONTROLS, INC.
3526 E FT LOWELL RD - PO BOX 40020 - TUCSON AZ 85717-0020
TEL: 520-881-2130 FAX 520-881-4983

CERTIFICATION OF CONFORMANCE

Customer: California Institute of Tech
Purchase Order: 75ADV-S109097
Quantity shipped: 20
Description: HUMIDITY INDICATOR
AGM Cage & Part Number: 08992 TA370-234P
Customer Number: 19960050
Sales Order: S201101448 3 02-11-2011

THIS IS TO CERTIFY that all materials, parts, and/or processes used in the manufacture of articles on this order are in accordance with the requirements, applicable drawings, and/or specifications called for on the purchase order referenced above.

We further certify that all products in this order are free of solder, platings, coatings, and claddings that exhibit either of the following:

- a. Material composition greater than 97% tin by weight.
- b. Tin alloy with composition of less than 3% lead by weight.

Parts did not come into direct contact with mercury or any of its compounds nor with any mercury containing devices employing a single boundary or containment during the manufacturing processes, tests or inspection while at our facility. This product was not manufactured with and does not contain any ozone depleting substances.

AGM Container Controls is Certified and Registered to be in compliance with ISO 9001:2008.

Richard L. BeMiller
Quality Systems Manager

2-11-11
Date