



---

# SenseMonitor Updates for S3

Patrick Sutton  
LIGO-Caltech

# Outline of Talk

---

- SenseMonitor review.
- SenseMonitor updates for S3.
- Preliminary S3 performance.

# SenseMonitor

- Estimates average range to which IFO can detect the inspiral of 1.4-1.4 $M_{\odot}$  neutron star binary:

$$\text{Range} = \left[ \frac{5\mathcal{M}^{5/3}\Theta^2}{96\pi^{4/3}\rho^2} \int_{f_l}^{f_h} df \frac{f^{-7/3}}{S(f)} \right]^{1/2}$$

↑
↑  
 constant                      noise power varies

- Major task is estimating strain noise spectrum (tracking the AS\_Q-to-strain calibration).
- Outputs range and calibration data to DMViewer, trend frames, web log files.

# SenseMonitor (cont'd)

- Calibrating the noise  $S(f)$ :



$$X_h(f) = \frac{1 + \alpha\beta G(f)}{\alpha C(f)} X_{AS-Q}(f)$$

- Reference calibrations (open-loop gain  $G$ , sensing function  $C$ ) measured during a calibration run. Fixed.
- $\alpha$  is optical gain change:
  - » varies over ~minutes
  - » get from amplitude of calibration line
- $\beta$  is DARM gain change
  - » fixed during S2, varies over ~minutes in S3
  - » get from DARM\_GAIN, ICMTRX\_01 channels

# Upgrade for S3: More Robust Calibration


---

- **Problem:** Injected amplitude of calibration line changed several times during S2 for H1, L1, causing unphysical changes in  $\alpha$ .
- **Solution:** SenseMonitor now monitors injection channel to determine amplitude at which line is injected.
  
- **Problem:** For S3 have time-dependent ICMTRX\_01 ( $\beta$ ).
- **Solution:** SenseMonitor now monitors relevant DARM channels to track  $\beta$ .




# Upgrade for S3: Additional Plots

- Calibrated noise spectra available in DMTViewer :



# Upgrade for S3: Additional Plots

- Range integrand  $f^{-7/3}/S_h(f)$  available in DMTViewer :







# S3 Status

---

- SenseMonitor running for all IFOs.
- LHO:
  - » Reference calibrations measured during E10.
  - » E10/S3 SenseMonitor range estimates for H1, H2 have been verified **independently** (few %) by Mike Landry's AutoCalibrator. (Only common assumptions are DC calibrations.)
- LLO:
  - » E10 references out of date due to ongoing commissioning; recently restarted using updated interim references (~25% range change). Validation ongoing.

# IFO performance since start of S3





• 25% - 50% - 75% percentile ranges (Mpc):  
H1: 1.71, 1.99, 2.32  
H2: 0.870, 0.910, 0.951  
L1: 0.692, 0.789, 0.871 [+ 0-50%]

# Summary

---

- SenseMonitor now tracks excitation and darm channels for more robust calibration tracking – less user intervention required.
- Independent range verification available from AutoCalibrator (LHO) and InspiralRange (LLO) tools.
- Calibrated noise spectra, range integrals sent to DMTViewer.
- Up-to-date documentation is available from the DMT spi page at the sites.